LABORATORY EQUIPMENTS







Note: The durometer presented is only for evaluation purposes and is not included on the bench.

Operating stand for "SHORE A, D" durometers



Standards: DIN 53 505; ISO 868

ZIPOR - Equipamentos e Tecnologia Industrial, S.A.

Commercial Services:

Tel.: +351.256 831 411 **Fax:** +351.256 831 412

email: pegasil@zipor.com

<u>Technical Services</u>:

Tel.: +351.256 888 240 **Fax:** +351.256 888 121

LABORATORY EQUIPMENTS

Operating stand for "SHORE A, D" durometers

Pur	oose:
-----	-------

Ensure that the durometers are pressed plane-parallel to the material under test with a standard known force.

Description:

Equipment main features:

- . Support basis with parallelism adjustment;
- . Handle to raise the test table;
- . Adjustable Durometer attachment system;
- . Stainless steel test table and attchment system;
- . Attachment system with height adjustment;
- . Suitable mass to exert the pressure required according to the hardness scale.

Supplied with:

Operating instructions; conformity / calibration certificate

Standards:

DIN 53 505; ISO 868

Power Supply: Power supply not required

Noise:

ZIPOR - Equipamentos e Tecnologia Industrial, S.A.

Commercial Services: Tel.: +351.256 831 411 Fax: +351.256 831 412

Power Consumption:

email: pegasil@zipor.com

<u>Technical Services</u>: Tel.: +351.256 888 240

Fax: +351.256 888 121